



NUMALIGARH REFINERY LIMITED

CORRIGENDUM - 2

Date: May- 15, 2019

[Total Pages 06]

TENDER NAME: Supply, Installation, Testing & Commissioning of Whole body multi slice C.T. Scanner capable of generating 32 slices or more per rotation, along with all accessories & with 05 (five) years warranty as per Annexure-1 at VKNRL Hospital.

TENDER No: OC09000217/SKD

With reference to the Notice Inviting Tender against the above mentioned tender please note the following changes after pre-bid meeting held on 26.04.2019 which shall be considered as a part of tender documents.

Commercial:

1. In AGREED TERMS & CONDITIONS (ATC) point no 12 PAYMENT TERMS of page no-05 of main tender document.

❖ **For Supply: 100% payment** through inland LC (Letter of Credit).

❖ **Out of 100% :**

- 80% payment against submission of documents through bank. Respective bank charges to the respective account. LC of 100% of total supply value shall be opened.
- Balance 20% payment only after successful completion of installation / erection & commissioning activities, submission of documentations etc. Subject to furnish PBG of 10% of the total order value valid up to Warranty period i.e. 05 years after completion of SITC job plus 03 months claim period.

2. In AGREED TERMS & CONDITIONS (ATC) point no 07 of page no-04 of main tender document may be read as follows:

07.	Modification (if any) of existing setup for installation of machine shall be in bidder's scope. Bidder may visit proposed site to assess the requirement of modification before quote.	Not Applicable
-----	--	-----------------------

3. In AGREED TERMS & CONDITIONS (ATC) point no 14 of page no-06 of main tender document may be read as follows:

14.	<u>DELIVERY PERIOD/CONTRACTUAL PERIOD:</u> Shall be 16 weeks from the date of PO/LOI. Bidder has to complete the supply, installation, testing, commissioning & handover the complete system within 16 weeks from the date of PO. For Supply part the date of LR shall be considered as date of delivery.	<u>DELIVERY PERIOD/CONTRACTUAL PERIOD:</u> Shall be 16 weeks from the date of opening of LC. Bidder has to complete the supply, installation, testing, commissioning & handover the complete system within 16 weeks from the date of opening of LC. For Supply part the date of LR shall be considered as date of delivery.
-----	---	---

4. **GST against supply part i.e. item no 10 as per BOQ:**

If it is not possible to quote single standard GST rate against item no 10 of BOQ because of multiple items with different GST rates, bidder may keep the GST rate against item 10 as “0%” It may indicate the same in % age of basic amount of item 10 [which may differ from standard GST rates 5%, 12%, 18% & 28%] in the **technical bid in ANNEXURE-G**. The total GST amount against item no 10 shall be derived from the quoted % age in **ANNEXURE-G** & quoted basic rate in BOQ for item no 10. The derived GST amount shall be considered at the time of Final price bid evaluation.

For example against item no 10 for 1 set there are total 05 items:

Item	Basic value	P&F @ 1%	Freight @2%	GST rate	GST amount
For main item	Rs 5000000.00	Rs 50000.00	Rs 100000.00	12%	Rs 618000.00
Assessories-1	Rs 500000.00	Rs 5000.00	Rs 10000.00	18%	Rs 92700.00
Assessories-2	Rs 200000.00	Rs 2000.00	Rs 4000.00	28%	Rs 57680.00
Assessories-3	Rs 100000.00	Rs 1000.00	Rs 2000.00	28%	Rs 28840.00
Assessories-4	Rs 50000.00	Rs 500.00	Rs 1000.00	5%	Rs 2575.00

Total basic: Rs 5850000.00
799795.00

Total GST amount: Rs

You have to enter against Basic price against 1 set as Rs 5850000.00 & GST % age as “0%” and indicate in the **technical bid in ANNEXURE-G** the total GST applicable against item no 10 as **13.671%** [799795/5850000] of basic quoted rate for item no 10 of BOQ.

After opening of price bid total GST amount against item no 10 shall be derived using the quoted % age in Technical bid & quoted price for item no 10 in the BOQ which shall be considered for final evaluation. **Do not mention any price in Annexure-G**

Successful bidder has to provide price break-up against total amount quoted for item no 10 corresponding to GST rates. Otherwise, bidder may quote the GST rate against item no 10

Technical: Refer Annexure 1 & changes as follows:

items	As per Tender	Modification [may be considered as follows.]
Technical Specification Page no 1 of Annexure-1	Whole body multi slice C.T. Scanner capable of generating 32 or more slices per rotation and system should be currently under production. The CT scanner should be latest and equipped with latest raw data based iterative reconstruction technology. System must be US FDA approved.	Whole body multi slice C.T. Scanner capable of generating 32 or more slices per rotation and system should be currently under production. The CT scanner should be latest and equipped with latest raw data based iterative reconstruction technology. System must be US FDA or CE (European) approved.
3. GANTRY AND SCANNING TABLE: point no 3 (C)	c. Scan Field of View (FOV)- 5cm to 40 cm. Please mention the FOV	c. Scan Field of View (FOV)- 5cm to 40 cm or more. Please mention the FOV
14. IMAGE ARCHIVING:	Image archiving on HD DVD / Blue Ray/ CD should be available. In case of CD-1000CDs may be supplied along with. USB connectivity should be provided	Image archiving on HD DVD / CD should be available. In case of CD-1000CDs should be supplied along with. USB connectivity is a must for archiving in external hard drive.
16. WORKSTATION: It should be standard. Ponit no 16 (f)	f) Dental CT with capability for life size images on film should be possible	f) Dental CT should be possible
16. WORKSTATION: It should be standard. Ponit no 16 (i)	i) CT colonoscopy with CAD capability, virtual dissection view and polyp analysis capability	i) Virtual CT colonoscopy should be possible.
OTHER CONSIDERATIONS - Page no 5 of Annexure-1	OTHER CONSIDERATIONS THAT MAY BE ADDED	OTHER CONSIDERATIONS THAT TO BE ADDED [MANDATORY]
	1. The system & the configurations mentioned above are mandatory and no items as asked above shall be treated as optional in the offer.	No change
	2. All compliance to the tender should be in the form of original data sheets or original certificate from the manufacturer. Items under work in progress will not be considered.	No change

	3. Onsite Application Training for the CT Technicians for a total period of 4 weeks.	No change
	4. Training for 1 Radiologist, 2 CT Technicians and 1 Bio-medical Engineer / IT Officer at a reputed institute within India for 1 week per person.	4. Training for 1 Radiologist, 2 CT Technicians and 1 Radiology Coordinator at a reputed institute within India for 1 week per person. Registration, cost of registration /training fees and co-ordination for the training will be bidder's responsibility. Travel and accommodation will be arranged by the VKNRL Hospital.
	5. Vendor shall have process the documents and get the clearance of the unit from AERB	No change
	6. Adequate supply of books and reference journals (one foreign journal within the warranty period) shall have to be provided.	6. Adequate supply of books and reference journals should be supplied, as per list provided by VK-NRL hospital. The books will have to be supplied within the time frame of installation & commissioning. Journals (including one foreign journal) shall have to be provided throughout the warranty period. Cost of books and journals will be within Rs.2,00,000/- (Rupees two lacs only).
	7. As there is continuous development of technology, the latest model available with the manufacturer shall be offered in the tender.	No change
	8. Radiation shielding for doors, walls, windows etc as per AERB norms.	8. Radiation shielding for doors, walls, windows etc. as per AERB norms should be specified. Implementation of the same shall be under scope of VKNRL Hospital scope

Other Terms & Conditions: Point no V page no 8 of Annexure-1	<u>DELIVERY PERIOD/CONTRACTUAL PERIOD:</u> Shall be 16 weeks from the date of PO. Bidder has to complete the supply, installation, testing, commissioning & handover the complete system within 16 weeks from the date of PO.	<u>DELIVERY PERIOD/CONTRACTUAL PERIOD:</u> Shall be 16 weeks from the date of opening of LC. Bidder has to complete the supply, installation, testing, commissioning & handover the complete system within 16 weeks from the date of opening of LC.
---	--	--

All the bidders are requested to upload this corrigendum along with the online offer with seal & sign.

All other terms and conditions of the tender document shall remain unchanged.

**GM (Commercial & Legal)-i/c
Numaligarh Refinery Limited,**

ANNEXURE-G

TENDER NAME: Supply, Installation, Testing & Commissioning of Whole body multi slice C.T. Scanner capable of generating 32 slices or more per rotation, along with all accessories & with 05 (five) years warranty as per Annexure-1 at VKNRL Hospital.

TENDER No: OC09000217/SKD

	<u>Bidder to Indicate[in %age only]</u>	<u>Remarks</u>
% age GST rate against item no 10 of BOQ [should be % age of basic amount quoted in the BOQ for item 10 for 1 set]		<u>Don't mention any price</u>

This mentioned % age shall be considered at the time of final price bid evaluation to derived the total GST amount against item no-10.

For example : If you quoted basic amount against item no 10 as Rs 1000.00 in the BOQ

& % age as 13.5% in the above table in the technical bid , total GST amount shall be Rs 135.00.